


Application 210-00

Agilent Reformulated Fuel Analyzer

Oxygenates in Gasoline and Naphtha
According to ASTM D5599 and EN1601

Technical Overview



Application Highlights


- ASTM Method D5599 and EN 1601
- Oxygenate Flame Ionization Detector (O-FID) to detect the following components to a lower detection limit of 0.06 wt%:

Methanol	Ethanol
Isopropanol	t-Butyl alcohol
n-Propanol	Methyl tert-butyl ether (MTBE)
Isobutyl alcohol	t-Amylmethyl ether (TAME)
n-Butanol	Diisopropyl ether (DIPE)
sec-Butanol	Ethyl tert-butyl ether (ETBE)
t-Pentanol	Dimethoxyethane

- Analysis time is approximately 15 minutes

Optional Configurations

ASTM D3606/ASTM D4815
ASTM D3606/ASTM D4815/ASTM D5580
ASTM D4815/ASTM D5580
ASTM D5769/ASTM D4815
ASTM D5769/ASTM D4815/ASTM D5580
ASTM D5769/ASTM D5599
ASTM D3606/ASTM D5580
ASTM D3606/ASTM D5599
ASTM D5769/ASTM D3606
ASTM D5769/ASTM D5580
ASTM D5769/ASTM D5599/ASTM D3606
ASTM D5769/ASTM D3606/ASTM D4815/ASTM D5580



For More Information


For more information on our products and services, visit our
Web site at www.agilent.com/chem.



Agilent Technologies



WASSON ECE
INSTRUMENTATION



O-FID output from the Agilent reformulated fuel analyzer.

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc. 2002

Printed in the USA
December 3, 2002
5988-6728EN