



## Eclipse Accessory Base Plate Instructions

**Part number:** 0010075400

Installation category I

Pollution degree 2

Safety class 3 (EN61010-1)

### Safety information

Before using this accessory, you must read the Safety Practices and Hazards section in your Cary Eclipse Hardware Manual.

A link to the Safety section can also be found on the main page of the Cary Eclipse online Help.

If the equipment is used in a manner not specified, the protection provided by the equipment may be impaired.

### Introduction

The Eclipse Accessory Base Plate is a multi-purpose accessory that allows you to mount any user customized accessory to the instrument.



**Figure 1.** Eclipse accessory base plate

### Installation

#### To install the Eclipse Accessory Base Plate:

- 1 Make sure that the sample compartment is empty.
- 2 Position the Eclipse Accessory Base Plate on the two locating pins at the rear of the sample compartment. The fastening lever should be at the front of the sample compartment.
- 3 Turn the fastening lever to the right to lock the Eclipse Accessory Base Plate into position.



**Agilent Technologies**

## Specifications

### Weights

Packed: 11.5 kg (25.3 lb)  
Unpacked: 8.0 kg (17.6 lb)

### Dimensions (W x D x H)

Packed: 385 x 700 x 360 mm (15.2 x 27.6 x 14.2 in)

## Maintenance and cleaning

All cleaning should be done with a soft cloth. If necessary, this cloth can be dampened with water or a mild detergent to clean black painted surfaces. Do not use organic solvents or abrasive cleaning agents.

This information is subject to change without notice.



8510183800

Part Number: 8510183800

Edition 01/11

Issue 2

Printed in Australia

© Agilent Technologies, Inc. 2000, 2011

Agilent Technologies

679 Springvale Road

Mulgrave, VIC 3170