

NOTICE: This document contains references to Varian. Please note that Varian, Inc. is now part of Agilent Technologies. For more information, go to www.agilent.com/chem.



Varian 240-MS GC/MS Ion Trap Mass Spectrometer Pre-installation Instructions

Checklist

NOTE: Do not unpack the boxes.

Place a check in the box after satisfying each requirement as described in the instructions. All requirements must be met before requesting installation.

NOTE: If the laboratory is not ready for installation when the Varian Representative arrives, Varian, Inc. reserves the right to invoice for the Representative's time.

Requirements	<input checked="" type="checkbox"/>
Principal installation site is in compliance with all relevant safety regulations.	<input type="checkbox"/>
User Representative will be available during the installation and certification period.	<input type="checkbox"/>
Entrance to the laboratory is at least 92 cm (36 in.) wide.	<input type="checkbox"/>
Sufficient bench space is available for all components.	<input type="checkbox"/>
Bench can support system weight.	<input type="checkbox"/>
Bench does not vibrate.	<input type="checkbox"/>
Laboratory temperature maintained between 16 and 30 °C.	<input type="checkbox"/>
Relative humidity maintained between 40 and 80%.	<input type="checkbox"/>
Laboratory is free of excessive particulate matter.	<input type="checkbox"/>
Ventilation system is suitable.	<input type="checkbox"/>
Specified electrical supply and power outlets are installed.	<input type="checkbox"/>
CI reagent gas (methane, isobutane, or ammonia: 99.99% pure), regulator, and gas lines are installed.	<input type="checkbox"/>
Helium (99.999% pure), regulator, and gas lines are installed.	<input type="checkbox"/>
Shipping cartons examined for damage. If there was any damage, the damaged shipping carton procedure was completed.	<input type="checkbox"/>

Request for Installation

After preparing your site, contact the Customer Service office in your region to schedule installation.

Contents

Introduction	2	Installation Site Power	11
Safety	2	Qualified Computer Equipment	15
Before the Installation	3	Gas Requirements	15
User Representative	3	Chemical Ionization Reagent Gases	15
Entrance	3	GC Carrier Gases	16
Bench Space and Load	3	When the GC/MS Arrives	17
Vibration	8	Inspection	17
Temperature	8	Unpacking and Installing	18
Humidity	8	Spare Parts	18
Particulate Matter	8	Preventive Maintenance	18
Ventilation System	9	Trademark Acknowledgment	18
Power Requirements	9		

Introduction

The Pre-installation Instructions guide you through each requirement of the checklist. Follow these instructions to ensure that the installation requires no more than the usual three days. Have the completed checklist available when you call to schedule the installation. After the Varian Representative has completed the installation, you can analyze your samples.

This document includes pre-installation instructions for the 431-GC and the 450-GC.

The GC/MS operates reliably under carefully controlled environmental conditions. It is your responsibility to provide suitable, power sources, operating environment and materials. Using or maintaining a system outside of the power and operating environment ranges and limits described in these instructions, may cause failures. The Warranty and Service contract specifically excludes the repair of failures due to such causes.



All phases of the installation site preparation must conform to local safety, electrical, and building codes. These codes take precedence over any recommendations in these instructions, and the customer is responsible for compliance.

Safety

Safety is the most important consideration for instrument use. Determine if the installation site complies with all relevant safety regulations.



Check the checklist box: *Principal installation site is in compliance with all relevant safety regulations.*

Before the Installation

User Representative

Schedule the installation when the User Representative is available, because one of the important duties of the Varian Representative is to familiarize the User Representative with the basic functions of the 240-MS GC/MS.

- Check the checklist box: *User Representative will be available during the installation and Varian certification period.*

Entrance

Before arranging for delivery of the GC/MS, determine that there is sufficient clearance to move the shipping container to the installation site. The shipping box is 72 cm (28 in.) wide by 92 cm (36 in.) long. If you move the instrument using the pallet, you need at least 92 cm (36 in.) clearance in the width of doorways. Allow additional room for maneuvering the shipping containers around corners or through doors.



CAUTION

The MS, foreline pump, and GC are heavy. To prevent personal injury, use appropriate moving and lifting techniques.

- Check the checklist box: *Entrance to the laboratory is at least 92 cm (36 in.) wide.*

Bench Space and Load

Use the following information to plan the layout for your system. The Varian Representative will unpack the boxes and place the modules on the bench. Figure 1 shows a possible layout for the 240-MS with the 431-GC. Figure 2 shows a possible layout for the 240-MS with the 450-GC.

240-MS with the 431-GC

In Figure 1, the 431-GC is to the left of the MS to allow the transfer line to connect from the GC to the 240-MS. The bench must be long enough and strong enough to support the weight of the system and any additional equipment, see Table 1 and Table 2. The bench must be at least 84 cm (33 in.) deep.



Figure 1 Layout of the 240-MS with the 431-GC

Table 1 Bench Space

Bench Length	Monitor and Keyboard	CPU	431-GC	240-MS	Total
cm	43.8	18.6	33	38	133.4 cm
in.	17.25	7.4	13	15	52.65 in.

Table 2 Bench Load

Bench Load	Monitor, CPU And Keyboard	431-GC	240-MS	Total
kg	17.9	23	42	82.9 kg
lb	39.4	51	93	183.4 lb

240-MS with the 450-GC

In Figure 1, the 450-GC is to the left of the MS to allow the transfer line to connect from the GC to the 240-MS. The bench must be long enough and strong enough to support the weight of the system and any additional equipment, see Table 3 and Table 4. The bench must be at least 84 cm (33 in.) deep.



Figure 2 Layout for the 240-MS with the 450-GC

Table 3 Bench Space of the 240-MS with the 450-GC

Bench Length	Monitor and Keyboard	CPU	450-GC	240-MS	Total
cm	43.8	18.6	66	38	166.4 cm
in.	17.25	7.4	26	15	65.65 in.

Table 4 Bench Load of the 240-MS with the 450-GC

Bench Load	Monitor, CPU and Keyboard	450-GC	240-MS	Total
kg	17.9	43	42	102.9 kg
lb	39.4	95	93	227.4 lb

Determining Space and Load

The 240-MS GC/MS, except for the foreline pump, belongs on a clean, flat bench. The foreline pump belongs under the bench. The vacuum line connecting the MS to the foreline pump is 1.2 m long (48 in.) long. The MS ships with a 1.7 m (66 in.) vacuum line attached. To accommodate the pump vacuum line, the bench should be no higher than 91 cm (36 in.). If your bench is higher, place a bench that can isolate vibrations under the pump. The bench must support the 22 kg (48 lb) pump. If your bench abuts a wall, drill a 3.8 cm (1.5 in.) diameter hole through the

rear of the bench for the vacuum hose.

Put the PC either on the same bench as the MS or on a separate table. If you put it on a separate table, position it within 3 m (10 ft) of the rear of the MS to accommodate the length of the connections. Table 5 has more information.

Table 5 Dimensions and Weights

Instrument	Height		Width		Depth		Weight	
	cm	in.	cm	in.	cm	in.	kg	lb
240-MS	46	18	38	15	66	26	42	93
431-GC	46	18	33	13	56	22	23	51
450-GC	53	21	66	26	56	22	43	95
Monitor	36.2	14.3	41	16.1	18.3	7.2	5.7	12.6
CPU	41	16	18.6	7.5	44.5	17.5	11.7	25.8

Allow bench space for other functions see Table 6.

Table 6 Bench Space Allowances

Purpose	Recommended Space Allowance
Provide access to the transfer line.	Allow at least 30 cm (12 in.) to the left side of the GC/MS.
Provide space for air circulation, gas lines, and electrical connections.	Allow at least 15 to 30 cm (6 to 12 in.) behind the GC/MS.
Dissipate heat and allow for routine maintenance.	Allow at least 76 cm (30 in.) above the GC/MS.

The bench must be at least 84 cm (33 in.) deep to accommodate the system. Use the Calculation Worksheets, Table 7 and

Table 8, to plan your system.

Table 7 Bench Space Calculation Worksheet for Your System

Bench Length	Monitor and Keyboard	CPU	_____ GC	MS	Other Components	Total
cm	43.8	18.6	_____	38	_____	_____ cm
in.	17.25	7.5	_____	15	_____	_____ in.

Table 8 Bench Load Calculation Worksheet for Your System

Bench Load	Monitor, Keyboard and CPU	_____ GC	MS	Other Components	Total
kg	17.9	_____	42	_____	_____ cm
lb	39.4	_____	93	_____	_____ in.



Check the checklist box: Sufficient bench space is available for all components.



Check the checklist box: *Bench can support system weight.*

Vibration

Ensure that the bench for the 240-MS GC/MS is free from vibrations, especially those caused by equipment in adjoining locations. Because the foreline pump vibrates during operation, put it on the floor below the MS, not alongside the system on the bench.



Check the checklist box: *Bench does not vibrate.*

Temperature

The optimal operating temperature is between 16° and 30 °C (61 and 86 °F).

NOTE: As laboratory temperature increases, system reliability decreases due to heat generated by electronic components during instrument use. This heat must dissipate to the surrounding air for reliable operation.

The airflow around the system must be adequate. The air conditioning system must be capable of maintaining a constant temperature in the immediate vicinity of the system. Do not place the system near air ducts, windows, or heating and cooling systems. The average steady-state heat load of the 240-MS GC/MS is 6,000 Btu, with a possible short-term heat dissipation of 15,000 Btu during startup.

Hot air vented from GC column ovens may contribute to room heating and to the resulting air conditioning load. Ducting the GC column oven air out of the laboratory reduces this heating effect.



Check the checklist box: *Laboratory temperature maintained between 16 and 30 °C.*

Humidity

The relative humidity of the operating environment must be between 40 and 80%, with no condensation. Operating the 240-MS GC/MS at a very low humidity may result in the accumulation and discharge of static electricity, shortening the life of electronic components. Operating the system at high humidity may produce condensation and result in short circuits. Put a temperature/humidity monitor in your laboratory.



Check the checklist box: *Relative humidity maintained between 40 and 80%.*

Particulate Matter

The laboratory must not have excessive dust, smoke, or other particulate matter. Particulate matter may block airflow vents causing the electronics to over heat.



Check the checklist box: *Laboratory is free of excessive particulate matter.*

Ventilation System

The foreline pump exhausts most compounds introduced into the MS along with oil vapor from the pump. Check that the ventilation system is suitable for the foreline pump. Consult local regulations.



Check the checklist box: *Ventilation system is suitable.*

Power Requirements

431-GC

The 431-GC requires a separate circuit and the outlet must have adequate amperage capacity and a reliable ground.

All voltages of the 431-GC are supplied with a standard molded power cord with an IEC60320 “Hot” connector. This power cord can be replaced to suit local power requirements.

Each 431-GC requires a power source capable of providing up to:

- 101 V ac \pm 10%, 50 or 60 Hz \pm 2%, 14 Amps, 1.4 Kilowatts
- 120 V ac \pm 10%, 60 Hz \pm 2%, 11 Amps, 1.3 Kilowatts
- 230 V ac \pm 10%, 50 Hz \pm 2%, 6 Amps, 1.3 Kilowatts

Installation Category: II (per Standard IEC664)

Power supply interruptions/Dips conform to EN61000-4-11: 1994

A voltage drop of 40% maximum, in a timeframe of 20 msec is allowed.

450-GC

The 450-GC requires a separate circuit and the outlet must have adequate amperage capacity and a reliable ground.

In 230V, 50 Hz countries, the 450-GC is supplied with a standard molded power cord with an IEC60320 “Hot” connector. This connector plugs into the receptacle on the rear of the GC. This power cord can be replaced to suit local power requirements. Power cords for North America and other 120V, 60 Hz applications are terminated in a 3-prong plug that requires a matching 120 V ac receptacle, see the NEMA 5-20P power plug and outlet in




Figure 3.

Each 450-GC requires a clean power source capable of providing up to:

- 101 V ac \pm 10%, 50 or 60 Hz \pm 2%, 25 Amps, 2.5 Kilowatts
- 120 V ac \pm 10%, 60 Hz \pm 2%, 20 Amps, 2.4 Kilowatts
- 230 V ac \pm 10%, 50 Hz \pm 2%, 10 Amps, 2.3 Kilowatts

Installation Category: II (per Standard IEC664).

Power supply interruptions/Dips conform to EN61000-4-11: 1994.

A voltage drop of 40% maximum, in a timeframe of 20 msec is allowed.

240-MS

The MS requires a separate circuit and the outlet must have adequate amperage capacity and a reliable ground.

The MS requires a clean power source capable of providing up to:

- 90-132V ac, 60 Hz \pm 3 Hz, 12 A, 1.5 kW
- 180-264V ac, 50 Hz \pm 3 Hz, 6 A, 1.5 kW

240-MS GC/MS

The 240-MS GC/MS requires the following:

- One dedicated duplex single-phase power source with earth grounds hard-wired to the main power panel ground for the GC.
- One dedicated fourplex power source for the MS, computer, and monitor.
- Additional sample preparation devices or test equipment requires a dedicated power.
- Never plug the MS and the GC into the same power source or the power source may overload. Never use the free outlet on any of the power sources for equipment that draws more than 2A.
- Within North America, these power sources must be 20A, 100-120V ac 60 Hz \pm 3 Hz.
- Outside North America, power sources must be 10A, 200-240V ac, 50 Hz \pm 3 Hz.

NOTE: 120V and 101V systems are designed for single phase use only.

Installation Site Power

The power supply of the installation site must be either 100-120V ac or 200-240 V ac, and be stable (free of fluctuations due to slow changes in the average voltage or to changes resulting from surges, sags, or transients). The voltage must meet IEC 1000-4-5 and IEC 1000-4-11 standards for voltage stability.

NOTE: If the quality of the power is a concern, use an uninterruptible power supply or a power conditioner, or both.

Table 9 has the current requirements for instruments and components of the 240-MS Ion Trap GC/MS Mass Spectrometer

Table 9 GC/MS Power Requirements

Instrument or Component	Max Current Draw (AMPS)	
	100-120V	200-240V
431-GC	11	6
450-GC	20	10
240-MS	10	5
Computer	3	1.5
Monitor	3	1.5
Printer	3	1.5

The power cable from the GC is approximately 2 m (6.7 ft) long and has a National Electronics Manufacturers Association (NEMA) 5-20P power plug, shown in




Figure 3. NEMA 5-20P plugs are rated at 20A and 120 V ac, see Figure 3.

The power cable from the 240-MS is approximately 2.5 m (8 ft) long and has a National Electronics Manufacturers Association (NEMA) 5-15P power plug shown in




Figure 3 NEMA 5-15P plugs are rated at 15A and 120V ac.

The power cables for the computer, monitor, and printer are approximately 2 m (7 ft) long. They have NEMA 5-15P plugs.

Systems shipped outside the United States and Canada have CEE 7/7 plugs. These plugs are rated at 16A and 230V ac.



Replacing or substituting power cords or plugs must be done with strict compliance with all regulations, including electrical codes, power cord color coding, and appropriate regulatory agency certification marks.




Figure 3 NEMA 5-20P, CEE 7/7, and NEMA 5-15P Outlets and Power Plugs

- Check the checklist box: **Specified electrical supply and power outlets installed.**

Qualified Computer Equipment

If you need to install the Varian MS Workstation software on a computer not purchased from Varian, you must ensure that the computer is adequately equipped and compatible with the operation of data system and its communication interfaces of the data system. Consult the current list of requirements, available at this web site.

http://www.varianinc.com/cgi-bin/nav?products/chrom/gcms/msws_computer_req

The Varian Representative uses only a Varian qualified computer when testing the system specifications. Varian does not guarantee the function of the Varian MS Workstation software on other computer hardware or operating systems.

NOTE: Contact your Sales Representative for a list of the currently qualified equipment or more information.

Gas Requirements

Chemical Ionization Reagent Gases

The Chemical Ionization (CI) mode uses methane, isobutane, or ammonia as a reagent gas. CI operation consumes about 1 to 2 mL/min of reagent gas. The CI reagent gas should contain less than 1 ppm water. The reagent gas supply line connects directly to the CI gas inlet using 1/8 inch Swagelok® fittings.

New gas lines typically contain significant amounts of adsorbed water vapor. Pump air through the gas line to evacuate the water. Use the shortest gas line possible to minimize pumping time.

Methane 99.99% purity

- Use a two-stage, 0-15 psi (0-1 bar) pressure regulator with a stainless steel diaphragm.
- Set the output pressure to the MS between 3 and 5 psi (0.2-0.3 bar).
- Use stainless steel gas lines for methane.
- Flame treat or solvent wash gas lines to remove oil.

Isobutane 99.99% purity

- Use a two-stage, 0-15 psi (0-1 bar) pressure regulator with a stainless steel diaphragm.
- Set the output pressure to the MS between 3 and 5 psi (0.2-0.3 bar).
- Use stainless steel gas lines for isobutane.
- Flame treat or solvent wash gas lines to remove oil.

Ammonia 99.99% purity

- Use a two-stage, 0-15 psi (0-1 bar) pressure regulator with a stainless steel diaphragm.
- Set the output pressure to the MS between 3 and 5 psi (0.2-0.3 bar).
- Use stainless steel gas lines for ammonia.
- Flame treat or solvent wash gas lines to remove oil.



**WARNING:
INHALATION
HAZARD**



**WARNING:
CHEMICAL HAZARD**

Ammonia, Anhydrous (CAS Number 7664-41-7)

Anhydrous Ammonia gas is highly toxic. Inhalation may cause irritation to eyes and throat and may cause pulmonary edema, which can result in serious injury or death. Repeated exposure to Anhydrous Ammonia may cause permanent lung damage.

- Refer to the Material Safety Data Sheet (MSDS) for Anhydrous Ammonia for exposure control/personal protection, handling/storage, accidental release, first aid, and fire fighting measure requirements.
- Only use anhydrous ammonia with an appropriate ventilation system for the instrument, the exhaust from the foreline pump, and the gas cylinder. Use appropriate safety shutoff valves for the anhydrous ammonia supply and interconnecting lines.

You are responsible for determining and implementing appropriate precautions when using anhydrous ammonia and for compliance with all governmental regulations. You are responsible for understanding and adhering to all safe laboratory practices concerning the use of toxic gases, including anhydrous ammonia.



Check the checklist box: *CI reagent gas, regulator, and gas lines are installed.*

GC Carrier Gases

Helium

Helium is required as a carrier gas for the GC, with a minimum of 99.999% ultra-high purity, and less than 1.0 ppm each of water, oxygen, and total hydrocarbons. The minimum requirement is one 257 ft³ tank with an Alltech regulator #AL8111, or equivalent tank. Use a two-stage 0-100 psi pressure regulator with a stainless steel diaphragm.



Check the checklist box: *Helium is at least 99.999% pure and the appropriate regulator and gas lines are installed.*



CAUTION

A carrier gas supply that has greater than 1 ppm oxygen or water may significantly affect the performance of the GC/MS. It may also damage components such as the capillary column, filaments, and multiplier. Verify that your gas suppliers use controlled tanks to ensure that purity standards are maintained. If you purchase pure gases in contaminated tanks, your system may become contaminated and require costly and time consuming repairs.

Optional Carrier Gas Filter

Use the optional carrier gas filters to protect your system from potential contamination and to ensure optimum performance. For replacement procedures, refer to the instructions enclosed with the filters. The easy-to-remove gas filter cartridge combines three highly adsorptive materials in one filter to remove water, oxygen, and organic compounds. Install the carrier gas filter so that the indicator is visible. Replace the filter when the indicator shows that the filter is saturated, or after one year of service, depending on which comes first. For replacement procedures, refer to the instructions enclosed with the filter.

When the GC/MS Arrives

Inspection

After the 240-MS GC/MS arrives, carefully inspect the exterior of the shipping cartons for evidence of any damage that might have occurred during shipment. Inspect the cartons for the following:

- Water stains.
- Cuts, punctures, or deep indentations.
- Crushed corners or excessively abraded edges.
- Blue beads in the Tip (N) Tell arrow point.

Two Tip (N) Tell indicators and labels are affixed to the exterior of the shipping boxes. Read and follow the instructions on the label. If the Tip (N) Tell arrow point is blue, the box was on its side or tipped in transit, and the instrument may be damaged.



If no external damage is apparent, sign the receiving documents, "*Received but not inspected*" to indicate that the boxes have not been opened.

Varian will not accept liability for damage if you received obviously damaged materials without noting the damage on the receiving documents.

Do not open any boxes. The Varian Representative opens them during installation. Move the shipping containers to a warm, dry, secure area near the installation site.

If a shipping carton shows evidence of damage, do the following damaged shipping carton procedure:

1. Report the conditions to the carrier when you receive the shipment.
2. Note the damage on all copies of the shipping documents.
3. Write a brief description of the damage.
4. Ask the driver to sign next to your comments to signify agreement with the observations.
5. Contact the appropriate Varian office to report the damage.

Systems are shipped either **FOB Varian** or **FOB Destination**. The manner of shipment determines who is responsible for filing a claim against the carrier if the system was damaged in transit. Most systems are shipped **FOB Varian**, so the purchaser and the carrier are responsible for any damages incurred during shipping. Contact the Varian office for assistance with filing claims and billing repairs. If the system is shipped **FOB Destination**, contact the Varian office, and that office will file a claim against the carrier.



Check the checklist box: *Shipping cartons examined for damage. If there was any damage, the damaged shipping carton procedure was completed.*

Unpacking and Installing

The Varian Representative will contact you to review the Pre-installation Checklist to ensure that you have satisfied all of the site requirements. The Varian Representative will unpack and install the instrument and demonstrate fundamental operation and maintenance procedures. The User Representative must be available during the installation.

The Varian Representative will demonstrate that your system meets the performance specifications unless there are additional criteria explicitly written into your sales contract.

Plan to analyze your samples only after the installation, which usually takes three days, is finished.

Spare Parts

The *240-MS Ion Trap Hardware Operation Manual* (part number 395416700) lists the spare parts for routine operation.

Preventive Maintenance

You are responsible for performing routine and preventive maintenance of the GC, MS, data system, and any other modules. Any instrument problems resulting from a contaminated gas supply are billable and not included in the Warranty.

Perform regular preventive maintenance to increase the life of the system, to maximize system operational time, and to optimize system performance. Refer to the *240-MS Ion Trap Hardware Operation Manual* for details. Your Varian Representative will describe and demonstrate these procedures during the installation.

Trademark Acknowledgment

Microsoft® and Windows® are registered trademarks of Microsoft Corporation. Swagelok® is a registered trademark of Swagelok Company, Solon, Ohio. Other brand and product names are trademarks or registered trademarks of their respective holders.