

Improving Productivity and Extending Column Life with Backflush

Application Brief

Chin-Kai Meng

All Industries

A previous application note [1] has shown that multiple GC signals and MS signals can be acquired from a single sample injection. When a 3-way splitter is connected to the end of a column, column effluent can be directed proportionally to two GC detectors as well as the MSD. This multi-signal configuration provides full-scan data for library searching, SIM data for quantitation, and element selective detector data for excellent selectivity and sensitivity from complex matrices.

The system used in this study consists of a 7683ALS, a 7890A GC with split/splitless inlet, 3-way splitter, μ ECD, dual flame photometric detector (DFPD), and a 5975C MSD. Figure 1 shows four chromatograms from a single injection of a milk extract. The synchronous SIM/scan feature of the 5975C MSD provides data useful for both screening (full scan data) and quantitation (SIM data). DFPD provides both P and S signals without the need to switch light filters.

Noticeably in the full scan TIC in Figure 1, a significant number of matrix peaks were observed after 32 minutes. It is not uncommon to add a “bake-out” oven ramp to clean the column after analyzing complex samples. The bake-out period is used to quickly push the late eluters out of the column to be ready for the next injection. Therefore, it is common to use a higher oven temperature than required for the analysis and an extended bake-out period at the end of a normal

Highlights

- Backflush – a simple technique to remove high boilers from the column faster and at a lower column temperature to cut down analysis time and increase column lifetime.
- The milk extract example shows that a 7-minute 280 °C backflush cleaned the column as well as a 33-minute 320 °C bake-out. The cycle time was reduced by more than 30%.
- Using backflush, excess column bleed and heavy residues will not be introduced into the MSD, thus reducing ion source contamination.




Figure 1. Four chromatograms collected simultaneously from a single injection of a milk extract.



Agilent Technologies

over program to clean out the column, which adds to the cycle time and shortens the column lifetime. Adding the bake-out period to the milk extract analysis, additional matrix peaks were observed even up to 72 minutes, while target compounds already eluted before 42 minutes. This means that 30 minutes were lost in productivity for each injection.

Backflush [2] is a simple technique to drastically decrease the cycle time by reversing the column flow to push the late eluters out of the inlet end of the column. Late eluters stay near the front of the column until the oven temperature is high enough to move them through the column. When the column flow is reversed before the late eluters start to move down the column, these late eluters will take less time and at a lower oven temperature to exit the inlet end of the column.

There are many benefits in using backflush:

- Cycle time is reduced (no bake-out period, cooling down from a lower oven temperature)
- Column bleed is reduced (no high-temperature bake-out needed), resulting longer column life
- Ghost peaks are eliminated (no high boilers carryover into subsequent runs)
- Contamination that goes into the detector is minimized, which is especially valuable for the MSD (less ion source cleaning)

Figure 2 shows three total ion chromatograms from the Agilent 7890A GC/5975C MSD. The top chromatogram is a milk extract analysis with all the target compounds eluted before 42 minutes (over program goes to 280 °C). However, an additional 33-minute bake-out period at 320 °C was needed to move the high boilers out of the column. This bake-out period was almost as long as the required time to elute all target compounds. The middle chromatogram is the same milk extract analysis stopped at 42 minutes with a 7-minute backflush post-run at 280 °C added to the analysis. The bottom chromatogram is a blank run after the backflushing was completed. The blank run shows that the column was very clean after backflushing. The example shows that a 7-minute backflush cleaned the column as well as a 33-minute bake-out.

The milk extract example in Figure 2 illustrates the backflush technique in reducing cycle time and column bleed. The cycle time was reduced by more than 30% and the column was kept at 280 °C, without going to the bake-out temperature




Figure 2. Three total ion chromatograms comparing the results with and without backflush.

of 320 °C. A column effluent splitter or QuickSwap is required to do the backflush.

References

1. Chin-Kai Meng and Bruce Quimby, "Identifying Pesticides with Full Scan, SIM, μ ECD, and FPD from a Single Injection," Agilent Application Note, 5989-3299EN, July 2005.
2. Matthew Klee, "Simplified Backflush Using Agilent 6890 GC Post Run Command," Agilent Application Note, 5989-5111EN, June 2006.

Acknowledgement

Milk extract is courtesy of Dr. Steven Lehotay from USDA Agricultural Research Service in Wyndmoor, Pennsylvania, USA.

For More Information

For more information on our products and services, visit our Web site at www.agilent.com/chem.

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc. 2006

Printed in the USA
December 26, 2006
5989-6018EN

