



Restek Refined

ASTM Petrochemical Method Chromatography Product Guide

Proven, Integrated Solutions and Veteran
Expertise for Your Petrochemical &
Chemical Analyses



RESTEK

Pure Chromatography

www.restek.com

ASTM Petrochemical Method Chromatography Product Guide

Restek is your ideal partner for integrated petrochemical solutions, and the following ASTM method product guide will help you quickly pick the right GC columns and reference standards for SimDist, DHA, finished gasoline, and other common petroleum analyses.

If you have any questions or need more information, visit www.restek.com/petro for additional resources or to contact one of our in-house petroleum experts for assistance.

(High-Temperature) Simulated Distillation (SimDist)

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D2887	Standard Test Method for Boiling Range Distribution of Petroleum Fractions by Gas Chromatography (C5–C44)	MXT-2887, Siltek-treated stainless steel 10 m x 0.53 mm x 2.65 µm - cat.# 70199 or MXT-1HT SimDist, Siltek-treated stainless steel 10 m x 0.53 mm x 2.65 µm - cat.# 70132 or Rtx-2887 Column 10 m x 0.53 mm x 2.65 µm - cat.# 10199	ASTM D2887-12 Calibration Standards - cat.# 31674 Polywax Standards - cat.# 36224–36227 D2887 Calibration Mix - cat.# 31222	GC_PC1299 GC_PC00750
D2887B Accelerated	Standard Test Method for Boiling Range Distribution of Petroleum Fractions by Gas Chromatography (C5–C44)	MXT-1HT SimDist Column 10 m x 0.53 mm x 0.88 µm - cat.# 70134 or MXT-1HT SimDist Column 5 m x 0.53 mm x 2.65 µm - custom column #1396	ASTM D2887-12 Calibration Standards - cat.# 31674 Polywax Standards - cat.# 36224–36227 D2887 Calibration Mix - cat.# 31222	GC_PC1303
D7213	Standard Test Method for Boiling Range Distribution of Petroleum Distillates in the Boiling Range from 100 to 615 °C by Gas Chromatography (C5–C60)	MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.10 µm - cat.# 70112 or MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.20 µm - cat.# 70115 or MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.88 µm - cat.# 70131	Request today	GC_PC1285 GC_PC1191
D6352	Standard Test Method for Boiling Range Distribution of Petroleum Distillates in Boiling Range from 174 to 700 °C by Gas Chromatography (C10–C90)	MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.10 µm - cat.# 70112 or MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.20 µm - cat.# 70115	Polywax Standards - cat.# 36224–36227	GC_PC1188
D7398	Standard Test Method for Boiling Range Distribution of Fatty Acid Methyl Esters (FAME) in the Boiling Range from 100 to 615 °C by Gas Chromatography	MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.88 µm - cat.# 70131	Polywax Standards - cat.# 36224–36227	GC_PC1285
D7500	Standard Test Method for Determination of Boiling Range Distribution of Distillates and Lubricating Base Oils in Boiling Range from 100 to 735 °C by Gas Chromatography (C7–C110)	MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.10 µm - cat.# 70112 or MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.20 µm - cat.# 70115	Polywax Standards - cat.# 36224–36227	GC_PC1164 GC_PC1165
D7169	Standard Test Method for Boiling Point Distribution of Samples with Residues Such as Crude Oils and Atmospheric and Vacuum Residues by High-Temperature Gas Chromatography	MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.10 µm - cat.# 70112 or MXT-1HT SimDist, Siltek-treated stainless steel 5 m x 0.53 mm x 0.20 µm - cat.# 70115	Polywax Standards - cat.# 36224–36227	GC_PC1308
D7096 (replaces D3710)	Standard Test Method for Determination of the Boiling Range Distribution of Gasoline by Wide-Bore Capillary Gas Chromatography	MXT-1, Siltek-treated stainless steel 15 m x 0.53 mm x 5.00 µm - cat.# 70177 or MXT-1, Siltek-treated stainless steel 30 m x 0.53 mm x 5.00 µm - cat.# 70179	Request today or D3710-95 Calibration Mix - cat.# 31223	GC_PC1326 GC_PC1327 GC_PC1328

Detailed Hydrocarbon Analysis (DHA)

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D5134	Standard Test Method for Detailed Analysis of Petroleum Naphthas through <i>n</i> -Nonane by Capillary Gas Chromatography	Rtx-DHA-50 50 m x 0.20 mm x 0.50 µm - cat.# 10147	Request today	GC_PC1318 GC_PC1319 GC_PC1320
D6729	Standard Test Method for Determination of Individual Components in Spark Ignition Engine Fuels by 100-Meter Capillary High-Resolution Gas Chromatography	Rtx-DHA-100 100 m x 0.25 mm x 0.50 µm - cat.# 10148	DHA Standards - cat.# 33034 - cat.# 30725–30731	GC_PC1322 GC_PC1323 GC_PC1324 GC_PC1325 GC_PC1330
D6730	Standard Test Method for Determination of Individual Components in Spark Ignition Engine Fuels by 100-Meter Capillary (with precolumn) High-Resolution Gas Chromatography	Rtx-DHA-100 100 m x 0.25 mm x 0.50 µm - cat.# 10148 <i>and</i> Rtx-5 DHA Tuning 5 m x 0.25 mm x 1.00 µm - cat.# 10165	DHA Standards - cat.# 33034 - cat.# 30725–30731	GC_PC1284
D6733	Standard Test Method for Determination of Individual Components in Spark Ignition Engine Fuels by 50-Meter Capillary High-Resolution Gas Chromatography	Rtx-DHA-50 50 m x 0.20 mm x 0.50 µm - cat.# 10147	DHA Standards - cat.# 33034 - cat.# 30725–30731	

Finished Gasoline

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D3606	Standard Test Method for Determination of Benzene and Toluene in Finished Motor and Aviation Gasoline by Gas Chromatography	D3606 Application 2-Column Set - cat.# 83606 <i>Specified in the D3606 method addendum — includes:</i> - Rtx-1 6' (1.8 m), 1/8" OD, 2.0 mm ID <i>and</i> - proprietary packing 16' (4.9 m), 1/8" OD, 2.0 mm ID	D3606 Standards - cat.# 30647–30674	GC_PC01121 GC_PC01103
D4815	Standard Test Method for Determination of MTBE, ETBE, TAME, DIPE, tertiary-Amyl Alcohol, and C1 to C4 Alcohols in Gasoline by Gas Chromatography (Oxygenates)	Micropacked with 20% TCEP on 80/100 Chromosorb PAW 0.56 m x 0.75 mm ID x 1/16" OD - cat.# 19040 <i>and</i> Rtx-1 30 m x 0.53 mm x 3.00 µm - cat.# 10185	Request today	GC_PC00194
D5580	Standard Test Method for Determination of Benzene, Toluene, Ethylbenzene, <i>p/m</i> -Xylene, <i>o</i> -Xylene, C9 and Heavier Aromatics, and Total Aromatics in Finished Gasoline by Gas Chromatography	Micropacked with 20% TCEP on 80/100 Chromosorb PAW 0.56 m x 0.75 mm ID x 1/16" OD - cat.# 19040 <i>and</i> Rtx-1 30 m x 0.53 mm x 5.00 µm - cat.# 10179	Request today	
D5501	Standard Test Method for Determination of Ethanol Content of Denatured Fuel Ethanol by Gas Chromatography	Rtx-DHA-150 150 m x 0.25 mm x 1.00 µm - cat.# 10149 <i>or</i> Rtx-DHA-100 100 m x 0.25 mm x 0.50 µm - cat.# 10148	Request today	GC_PC1144

Biodiesel

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D6584	Test Method for Determination of Free and Total Glycerin in B-100 Biodiesel Methyl Esters by Gas Chromatography	MXT-Biodiesel TG, Siltek-treated stainless steel 14 m x 0.53 mm x 0.16 µm with 2 m Integra-Gap - cat.# 70289 <i>or</i> MXT-Biodiesel TG, Siltek-treated stainless steel 15 m x 0.32 mm x 0.10 µm with 2 m x 0.53 mm retention gap - cat.# 70291 <i>or</i> Rtx-Biodiesel TG 15 m x 0.32 mm x 0.10 µm with 2 m x 0.53 mm retention gap - cat.# 10293	Biodiesel Standards - cat.# 31880 - cat.# 33020–33026 - cat.# 33032–33033	GC_PC1293 GC_PC00969

Visit us at www.restek.com/petro

Natural Gas

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D1945	Standard Test Method for Analysis of Natural Gas by Gas Chromatography	MXT-Msieve 5A, Siltek-treated stainless steel 30 m x 0.53 mm x 50 µm - cat.# 79723-273 <i>and</i> MXT-Q-BOND, Siltek-treated stainless steel 30 m x 0.53 mm x 20 µm - cat.# 79716-273	Natural Gas Standards - cat.# 34438-34440	GC_PC1182 GC_PC00169

Refinery Gas

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D2163	Standard Test Method for Determination of Hydrocarbons in Liquefied Petroleum (LP) Gases and Propane/Propene Mixtures by Gas Chromatography	Rt-Alumina BOND/Na ₂ SO ₄ 50 m x 0.53 mm x 10 µm - cat.# 19756	Refinery Gas Standards - cat.# 34441-34443	GC_PC1309 GC_PC1310
UOP 539	Standard Practice for Analysis of Reformed Gas by Gas Chromatography	Zabc Refinery Gas Packed Column Set - cat.# 88000-875 <i>or</i> MXT-Msieve 5A, Siltek-treated stainless steel 30 m x 0.53 mm x 50 µm - cat.# 79723-273 <i>and</i> MXT-Q-BOND, Siltek-treated stainless steel 30 m x 0.53 mm x 20 µm - cat.# 79716	Request today	GC_PC1182 GC_PC00169

Impurities

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D2593	Standard Test Method for Butadiene Purity and Hydrocarbon Impurities by Gas Chromatography	Rt-Alumina BOND/MAPD 50 m x 0.53 mm x 10 µm - cat.# 19778	Refinery Gas Standard #5 - cat.# 34443	GC_PC1202
D2712	Standard Test Method for Hydrocarbon Traces in Propylene Concentrates by Gas Chromatography	Rt-Alumina BOND/Na ₂ SO ₄ 50 m x 0.53 mm x 10 µm - cat.# 19756	Refinery Gas Standard #5 - cat.# 34443	GC_PC01142
D6159	Standard Test Method for Determination of Hydrocarbon Impurities in Ethylene by Gas Chromatography	Rt-Alumina BOND/KCl 50 m x 0.53 mm x 10 µm - cat.# 19760 <i>and</i> Rtx-1 30 m x 0.53 mm x 5.00 µm - cat.# 10179	Refinery Gas Standard #5 - cat.# 34443	GC_PC01110
D5441	Standard Test Method for Analysis of Methyl Tert-Butyl Ether (MTBE) by Gas Chromatography	Rtx-DHA-100 100 m x 0.25 mm x 0.50 µm - cat.# 10148 <i>or</i> Rtx-DHA-50 50 m x 0.20 mm x 0.50 µm - cat.# 10147 <i>or</i> Rtx-DHA-150 150 m x 0.25 mm x 1.0 µm - cat.# 10149	Request today	GC_PC1268 GC_PC1270 GC_PC1269

Sulfur

Method #	Method Title	Restek Column(s)	Restek Reference Standard(s)	Chromatogram#
D6228	Standard Test Method for Determination of Sulfur Compounds in Natural Gas and Gaseous Fuels by Gas Chromatography and Flame Photometric Detection	Rtx-1 60 m x 0.53 mm x 7.00 µm - cat.# 10193 <i>or</i> MXT-1, Siltek-treated stainless steel 60 m x 0.53 mm x 7.00 µm - cat.# 70193 <i>or</i> Rtx-1 60 m x 0.32 mm x 5.0 µm - cat.# 10180	Request today	GC_PC1223
D5623	Standard Test Method for Sulfur Compounds in Light Petroleum Liquids by Gas Chromatography and Sulfur Selective Detection	Rtx-1 30 m x 0.32 mm x 4.00 µm - cat.# 10198	Request today (gas, liquid)	GC_PC1271 GC_PC00198

RESTEK
Pure Chromatography

Questions? Contact us or your local Restek representative (www.restek.com/contact-us).

Restek patents and trademarks are the property of Restek Corporation. (See www.restek.com/Patents-Trademarks for full list.) Other trademarks in Restek literature or on its website are the property of their respective owners. Restek registered trademarks are registered in the U.S. and may also be registered in other countries. To unsubscribe from future Restek communications or to update your preferences, visit www.restek.com/subscribe. To update your status with an authorized Restek distributor or instrument channel partner, please contact them directly.

© 2019 Restek Corporation. All rights reserved. Printed in the U.S.A.

www.restek.com



Lit. Cat.# PCAR2082D-UNV